Groping for Quantitative Digital 3-D *Image Analysis:* An Approach to Quantitative Fluorescence in situ Hybridization in Thick Tissue Sections of Prostate Carcinoma

Karsten Rodenacker Michaela Aubele Peter Hutzler
P. S. Umesh Adiga*
GSF National Research Center for Environment and Health
Institute of Pathology, Dept. Biomedical Image Analysis
Ingolstädter Landstr. 1
D-85764 Neuherberg, Deutschland
e-mail: rodena@gsf.de
Tel: +49(0)89 3187 3401 *Fax*: +49(0)89 3187 3349

*Guest scientist, CVPR Unit, Indian Statistical Institute, Calcutta, India

Abstract

In molecular pathology numerical chromosome aberrations have been found to be decisive for the malignancy as well as prognosis of tumors. The existence of such aberrations can be detected by interphase fluorescence *in situ* hybridization (FISH). The gain (trisomy) or loss (monosomy) of certain base sequences in the deoxyribonucleic acid (DNA) can be estimated by counting the number of FISH signals *per cell nucleus*. The quantitative evaluation of such events is a necessary condition for a prospective use in diagnostic pathology. To guarantee a reproducible and unbiased signal count per nucleus, a whole nucleus has to be inspected. To avoid occlusions of signals the cell nucleus has to be analyzed in three dimensions. Confocal laser scanning microscopy is the means to obtain series of optical thin sections from fluorescence stained or marked material to fulfill the conditions mentioned above.

In the reported approach thick sections from prostate carcinomas were processed, digitized by confocal laser scanning microscopy (CLSM) and inspected by a computer-aided system. The preparative methods are shortly described. Main emphasis is given to the methodical questions of computer-aided analysis of large data sets. Questions of adequate data display and automated analysis are also faced. A graphical user interface (GUI) has been designed and tested for interactive data inspection by pathologists. Several automated analysis steps can be performed for quantification.

However tumor material in contrast to isolated or cultured cells is even for visual inspection a difficult material. Up to now a fully automated digital image analysis of 3-D data is not in sight. A semi-automatic segmentation method is presented.

1 Introduction

Molecular Pathology:

Prostate cancer is currently a leading cause of deaths from malignancies in American and European men [11]. Owing improved methods of detection the amount of cases being reported has increased sharply. The clinical course of prostate cancer is often unpredictable, and various efforts have been undertaken to improve reliability of prognostic parameters. Morphological grading, staging, and tumor size are mainly used to assess the malignant potential of prostate cancer [6]. Besides these methods new techniques like nucleic acid fluorescence *in situ* hybridization (FISH) become increasingly important in diagnostic and research pathology. By FISH numerical chromosome aberrations e.g. gaines or losses of specific single chromosomes (e.g. trisomy or monosomy) can be detected. Several studies have shown [3, 7] that trisomy of chromosome 7 is found in malignant and non-malignant tumors of lung, kidney, brain, as well as in tumors of the prostate. Considering the biological significance of this alteration being still under discussion *Zitzelsberger et al.* [17] have shown that trisomy of chromosome 7 may be used as prognostic factor because of correlation with the development of lymph node metastases and with grade of differentiation. Therefore, an improved image analysis procedure for quantitative evaluation of FISH in sections of prostate cancer has been established to simplify its quantitative evaluation, to confirm its prognostic value as well as to correlate FISH data with clinical follow-up in the long run.

Image Acquisition, Processing and Analysis:

The seemingly simple task of signal counts becomes difficult by the condition that the count has to be done related to a single cell nucleus. As far as these nuclei are isolated it is easy to estimate the membership of a signal to a nucleus. However in most tumor tissues this relation is difficult or even sometimes impossible to determine. This is valid for visual inspection and even more for computer based analysis. It is a well known fact that segmentation is the most cumbersome task in every digital image analysis approach. Recently another graphical user interface for visual inspection was presented [2] using primarily volume rendering techniques for display. In our presented approach we tried to mimic the inspection behaviour of pathologists by mainly presenting sections, however volume rendering is possible too. The interactive inspection part is already used for a large comparison of thin and thick section FISH already reported in [1].

2 Material: *Preparation, Staining and Sampling*

Tissue sections: 15 μm thick paraffin sections from routinely processed prostate cancer tissue were deparaffinized and rehydrated.

DNA probes and probe labeling: α-satellite DNA probes specific for the centromeric region of chromosome 7 were generated from clone 680TA [17]. DNA probe was labeled by nick-translation with biotin-dUTP (Boehringer, Mannheim) according standard procedure [17].

Hybridization: The fluorescence *in situ* hybridization (FISH) was performed according to [1]. Briefly, the slides were treated with pronase E (0.05% in 2xSSC, Merck, Darmstadt, FRG) at 37 °C for 20 to 35 minutes. After denaturation (70% formamide/2xSSC) at 72 °C for 20 to 45 minutes, the slides were dehydrated in an ethanol series. The hybridization mix consisted of carrier DNA (hering sperm DNA, SIGMA), 55% formamide/1xSSC, 10% dextran sulphate and the labeled DNA probe. This mixture was applied to the slides under a glass cover slip and sealed with rubber cement. After overnight incubation at 37 °C post-hybridization washes were performed at 43 °C in three changes of 50% formamide/2xSSC, PN buffer [8], and 3% bovine serum albumin (BSA, SIGMA).

Detection and counter stain: The biotin-labelled centromere 7 probe was detected by streptavidin-fluorescein isothiocyanat (FITC) and biotinylated anti-streptavidin conjugates (Camon, Wiesbaden). The nuclei were counter stained with propidium iodide (PI, SIGMA) and mounted with an antifade solution as described by [4]. Staining procedure was performed according to [14].

Sampling: Representative tumor and non-tumor (as control) regions with unambiguous FISH signals and well-preserved nuclei morphology were selected for acquisition by CLSM.

3 Methods

The development and application of image analysis procedures are performed under the software package IDL (Interactive Data Language. Version 4.0.1, Research Systems, Inc., Boulder, CO, USA). As hardware development platform a Silicon Graphics work station INDY is used with 128 MB main memory, true color dispay, connected via local area network to other computers and PC's,

especially the PC attached to the laser scanning microscope (LSM 410, Zeiss, Oberkochen, BRD).

Beside mere memory problems with 3-D images of sizes about 15 MB pixels even the processing time increases considerably. E.g. even a RISC processor of MIPS 4400 type becomes slow for methods usually applied and well known in 2D image analysis.

Image Acquisition

Fluorescence images are scanned using a Confocal Laser Scanning Microscope (CLSM) Zeiss LSM 410. Essential setup features for the acquisition of FISH signals in tissue sections are as follows: Lens Zeiss PNF 100x, 1.3, zoom =2, realized by scanning unit). The scanned field of 62.5 x 62.5 μm is sampled to 256 x 256 pixels giving a pixel resolution of 0.25 μm in x and y direction. Excitation laser lines are selected according to the fluorochromes used. For propidium iodide (PI) used as DNA counter stain and FITC labeled signals both are excited by the Argon line 488 nm. The emission of both is measured simultaneously in two separate channels using a bandpass 515 - 565 nm for FITC and a longpass LP 590 nm for the PI channel respectively. The axial distance selected between two subsequent confocal images depends on the further evaluation strategy. If a spatial isotropic representation of the 3-D data is intended, the axial distance is identical to the lateral distance of the pixels, i.e. 0.25 μm. For a 16 μm thick section this results in a sequence of 64 scanned images. If only the sampling theorem should be met roughly a 0.5 μm distance is enough, keeping in mind that the confocal axial resolution of a lens with NA = 1.3 is about the wavelength, i.e. 0.6 μm. The FITC emission is associated with the green channel and the PI emission with the red channel of a RGB true color image. The image data are stored on disk in TIFF format. Scanning one field of view of a section with 16 μm thickness and using 0.25 μm distance between individual images results in an uncompressed data set of 12.58 MByte.

Image Display

Various methods of display of digital volumes are proposed in this quickly growing field of methodical developments [5]. However, there are still large requirements in research to solve the technical as well psychological problems of display of 3-D informations.

To allow an observation familiar to pathologists, a GUI has been designed (Fig.1) displaying the 3-D image as a heap of orthogonal sections with arbitrary

choice of the cutting planes in x, y and z. Beside the counter stain volume in red color channel and the signal in green, a second type of signal may be stored in blue color channel. Possible display by interactions with the GUI are (Fig.1, 2):

1. Orthogonal sections of the original digitized volume with counter stain in red (R) and signal in green (G).

2. Thresholded and segmented counter stain to improve recognition of nuclear borders with adjustable parameters.

3. Segmented signals with adjustable parameters

4. Histograms from all channels, counter stain (HIST_R), signal (HIST_G) and blue channel (HIST_B) (Fig.1).

5. Surface rendered counter stain images,(SLIC_R), signal (SLIC_G) and results of segmentation (SLIC_B) (Fig.3).

6. Gallery of interactively selected box around or semi-automatic segmented nuclei as a sequence of sections of sub-windows (Fig.4, 5).

7. Journaling of the sequence of selected boxes, rsp. labelled objects including the signal count and other features. A replay function for repeated analysis of analysed locations is carried out.

Other possible interactions for analysis are histogram based local thresholding, cell labelling, FISH signal count, global and local segmentation, cutting of cell nuclei by spheres or ellipses, active volumes for segmentation, 3-D morphological operations, etc.

Interactive volume processing

For interactive volume analysis the orthogonal display of the volume can be positioned in x, y and z direction and a box around the object of interest resized by dragging can be selected (Fig.4). By another mouse click the seleceted sub-volume is displayed in a seperate window as a gallery of sub-sections (Fig.5). On left side all sub-sections and on right side zoomed sub-sections are shown.

During display automatically the number of FISH signals inside the selected sub-volume is calculated and the mouse cursor jumps to the counting button with respective number (Fig.2b). Clicking at the appropriate button stores the box location, size and count. In the display field of orthogonal sections the box location is outlined for information of the already scanned places. The buttons

SAVE, LOAD, CLEAR, REDRAW, DRAW Count and PRINT Count allow storage and display of the count list (Fig.2b). Also old count list files can be read and redisplayed for further inspection (Fig.6).

Semi-automatical volume processing

Segmentation of nuclei The complete automatic segmentation of cell nuclei is still a distant dream. We have developed a seeded volume growing technique based on several size and shape constraints, to segment the cell nuclei and to automatically count the FISH signal per nuclei. After slight volume opening the image is subject to global segmentation. It is automatically thresholded on the basis of local histograms and the 3-D connected components of the resulting two level image are labelled. Cells which are out of focus and too complicated to segment are rejected from further evaluation. Also the cells which are at the border of the image are rejected since the completeness of such a cell nuclei can not be assertained. Cells which are touching each other are first selected for semi-automatic segmentation. The center of such a cell is selected by clicking the mouse over approximate centroid. This point will act as a seed and the seed is grown in all the directions based on similarity constraint and other shape constraints. The similarity constraint α is defined as $\alpha = mean(markedcellregion) \sim U(i,j,k)$, where $U(i,j,k)$ is the input image volume.

If the neighboring voxels of the seed or marked region are well with in the similarity constraint threshold then such voxels are marked as cell regions. The overshoot of the marked region into neighboring cells which touch each other is checked by the shape constraints and an adjustable ellipsoid surface which is fitted in such a way that the cell nucleus falls just within.

After separation of the touching cell nucleus by volume growing method, the image volume is subjected to a relabelling process to enable the newly segmented cells to have a distinct label (Fig.7). Segmented cell regions are selected in the signal channel for further processing.

Segmentation of FISH signals: Several different methods beginning from thresholding via linear transformations to non-linear ones, e.g. transformations derived from mathematical morphology [12] were applied. For the actual material with small extension in x-, y- and z-direction FISH signals in 3-D are represented by small grain like structures.

A top-hat transformation was considered best to enhance the FISH signals for segmentation. The top-hat transformation image is the difference between an original grey scale image and an opened version. The radius of opening denotes

the diameter of the top-hat. A threshold of top-hat can be considered as the height of the top-hat. All operations are performed as 3-D operations.

After thresholding the top-hat transformation the resulting (binary) volume is processed by the selection of one connected object after the other, using 3-D neighborhood. A grain is accepted as FISH signal after a size and shape criterion and its being located inside the labelled nucleus.

Featuring

The goal of quantitative FISH about pathological alterations is the count of gains and losses of chromosomes. The physiological detectable disomic nuclei (2 signals) in tumor tissue are not in the first place of interest. Of most importance is the *number* of appearing trisomies (3 signals), polysomies (> 3 signals) and monosomies (1 signal). The subjective computer-aided evaluation should equip us with data about the spatial distribution of signals in the nuclei as well as about the size and shape of nuclei in 3-D.

Interactive: During visual inspection of the volume the coordinates and size of each selected box with the subjectively estimated signals are stored accompanied by the FISH signal count. This count list is subject to further analysis either by pooling count lists of several volumes or by using the box locations as expert selected seed for semi-automatical analysis.

Semi-Automatical: First a global (threshold) based segmentation and object labelling is performed. The pathologist can display the result and decide which nuclei are to be counted and which are to be deleted. Selection or deletion are done by clicking at the labelled object. Improvement of segmentation is performed by selecting the object and proceeding as explained earlier under the section segmentation of nuclei. For labelled objects the FISH signals can be automatically counted, the box coordinates of the surrounding box, the count, signal and nucleus volume and the label number are stored. The FISH signals are counted of either all labelled nuclei or only of interactively selected ones. In the latter case a gallery of a box determined by the size of the selected object can be displayed, comparable to interactive processing.

4 Preliminary Results and Conclusions

The interactive evaluation aid described above as well as the semi- automated evaluation deliver us with data for a more straight forward method of detection

of suspected or to be analyzed signals. This would allow to concentrate analysis steps to small regions of interest in the volume data.

Up to now 19 tumors with about 300 image (volume) blocks are visually evaluated using this programm. The results are presented in [1]. The comparison of visually counted signals and semi-automatic counting will be communicated in near future.

The methodical proceedings for quantitative fluorescent *in situ* hybridization for 3-D evaluation is outlined. A program which can run on multiple platforms is developed. The software is made user friendly. The pathologists who are not familiar with the computer aided image analysis, can easily work with this program. The software is not just for FISH evaluation but can be used for morphological feature evaluation and to work on any kind of visualizable signals in the cell nucleus. The GUI can also be used to visualize the cell distribution and/or invasive pattern of the malignant cells in three dimensions.

References

[1] Aubele, M., Zitzelsberger, H., Szücs, S., Werner, M., Braselmann, H., Hutzler, P., Rodenacker, K., Lehmann, L., Minkus, G., Höfler, H.: Comparative FISH analysis of numerical chromosome 7 abnormalities in 5 μm and 15 μm paraffin-embedded tissue sections from prostatic carcinoma. Histochemistry (1997) in press

[2] Becker Jr., C.R.L., Mikel, U.V., Oliver, LTC W.R., Sesterhenn, I.A.: Enumeration of interphase chromosomes *Comparison of visual* in situ *hybridization and confocal fluorescence* in situ *hybridization* Anal. Quant. Cytol. Histol. **18** 5 (1996) 405–409

[3] Babu, V.R., Miles, B.J., Cerny, J.C., Weiss, L., van Dyke, D.L.: Cytogenetic study of four cancers of prostate. Cancer Genet. Cytogenet. **48** (1990) 83–87

[4] Johnson, G.D., de C Nogueira Araujo, G.M.: A simple method of reducing the fading of immunofluorescence during microscopy. Immunol Methods **43** (1981) 349–350

[5] Kriete, A. Ed.: Visualization in biomedical microscopies - 3-d imaging and computer applications. VCH, Weinheim (1992)

[6] McNeal, J.E.: Cancer volume and site of origin of adenocarcinoma of the prostate: relationship to local and distant spread. Hum. Pathol. **23** (1992) 258-266

[7] Micale, M.A., Mohamed, A., Sakr, W., Powell, I.J., Wolman, S.R.: Cytogenetics of primary adenocarcinoma. Cancer Genet Cytogenet **61** (1992) 165–173

[8] Pinkel, D., Straume, T., Gray, J.W.: Cytogenetic analysis using quantitative high-sensitivity fluorescence hybridization. Proc Natl Acad Sci USA **83** (1986) 2934–2938

[9] Preston Jr., K. and Siderits, R.: New technologies for 3-d data analysis in histopathology. Anal. Quant. Cytol. Histol. **14** 5 (1992) 398–406

[10] Rigaut, J.P., Carvajal-Gonzalez, S., Vassy, J.: 3-d image cytometry. In: [5], 205–248

[11] Scardino, P.T., Weaver, R., Hudson, M.A.: Early detection of prostate cancer. Hum. Pathol. **23** (1992) 211–222

[12] Serra, J.: Image analysis and mathematical morphology. Academic Press, London (1982)

[13] Sheppard, C.J.R and Gu, M.: 3-d transfer functions in confocal scanning microscopy. In: [5] 251–281

[14] Tekola, P., Baak, J.P.A., Beliën, J.A.M., Brugghe, J.: Highly sensitive, specific, and stable new fluorescent DNA stains for confocal laser microscopy and image processing of normal paraffin sections. Cytometry **17** (1994) 191–195

[15] Thompson, C.T., LeBoit, P.E., Nederlof, P.M., Gray, J.W.: Thich-section fluorescence *in situ* hybridization on formalin-fixed, paraffin-embedded archival tissue provides a histogenetic profile. Am. J. Pathol. **144** 2 (1994) 237–243

[16] Zhu, Q., Tekola, P., Baak, J.P.A., Beliën, J.A.M.: Measurement by confocal laser scanning microscopy of the volume of epidermal nuclei in thick skin sections. Anal. Quant. Cytol. Histol. **16** 2 (1994) 145–152

[17] Zitzelsberger, H., Szücs, S., Weier, H.-U., Lehmann, L., Braselmann, H., Enders, S., Schilling, A., Breul, J., Höfler, H., Bauchinger, M.: Numerical abnormalities of chromosome 7 in human prostate cancer detected by fluorescence *in situ* hybridization (FISH) on paraffin-embedded tissue sections with centromere-specific DNA probes. J. Pathol. **172** (1994) 325–335

List of Figures

1 Graphical user interface overview 12
2 Possible User interactions 13
3 Graphical user interface for volume rendering and slicing 14
4 Display of orthogonal sections 15
5 Gallery of a selected box 16
6 Result display 16
7 Labelled section 17



Figure 1: Graphical user interface for 3-D image display



Figure 2: a | b Functional tablet for user interaction, a) Segmentation and semi-automatic count and journaling b) Display, visual count and journaling



Figure 3: Graphical user interface for volume display by rendering and slicing



Figure 4: Orthogonal sections in x-y-z with outlined selection box



Figure 5: Gallery of a selected box



Figure 6: a | b Result display with a) marked objects and b) FISH signal distribution



Figure 7: a | b | c Labelled section, a) globally segmented, b) locally improved by cutting touching nuclei, c) cleaned by deletion